

BLANK PAGE



Indian Standard

MINIMUM REQUIREMENTS FOR GENERAL PURPOSE AUDIO FREQUENCY SIGNAL GENERATORS (30 c/s to 30 kc/s)

(First Reprint JULY 1984)

UDC 621.373.029.4:621.317.3



@ Copyright 1967

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG

Indian Standard

MINIMUM REQUIREMENTS FOR GENERAL PURPOSE AUDIO FREQUENCY SIGNAL GENERATORS

(30 c/s to 30 kc/s)

Electronic Equipment Sectional Committee, ETDC 24

Chairman

Representing

SHRI S. THIRUVENKATACHARI

Directorate General of All India Radio (Ministry of Information & Broadcasting)

Members

RESEARCH ENGINEER, AIR (Alternate to

Shri S. Thiruvenkatachari) SHRI K. M. BALCHANDANI

Overseas Communications Service (Department of Communications)

SHRI G. D. GOKARAN (Alternate)

SHRI M. K. BASU

Wireless Planning & Co-ordination Wing (Department of Communications)

DR SABOJ DATTA

Directorate General of Civil Aviation (Department of Civil Aviation) (Ministry of Aviation & Transport)

SHRI B. R. V. VARDAN (Alternate) SERI P. N. DEOBHARTA

Directorate General of Technical Development (Ministry of Supply, Technical Development & Materials Planning)

SHRI A. N. MUKHERJEE (Alternate) DEPUTY DIRECTOR OF STORES Naval Headquarters (TECH)

STAFF OFFICER ELECTRICAL ENGINEERING (STANDARD-

IZATION) (Alternate)

SHRI H. JOGA RAO

State Broadcasting Department, Government of Andhra Pradesh, Hyderabad

Bharat Electronics Ltd, Bangalore SHRI S. T. KAGALI

SHRI K. S. KELKAR (Alternate)
SHRI R. S. MUNDKUB (Alternate)

SHRI M. N. KEKRE Police Wireless (Ministry of Home Affairs)

(Continued on page 2)

INDIAN STANDARDS INSTITUTION MANAK BHAVAN. 9 BAHADUR SHAH ZAFAR MARG **NEW DELHI**

(Continued from page 1) Representing Members Directorate of Technical Development & Produc-WG CDR J. S. LAMBA tion (Air) (Ministry of Defence) WG CDR S. RAO (Alternate) Directorate General of Inspection (Ministry of COL M. L. MIDHA Defence) LT-COL D. SWAROOP (Alternate) Indian Telephone Industries Ltd. Bangalore SHRI H. J. MIRCHANDANI SHBI S. D. MANI (Alternate) All India Radio Merchants' Association. Bombay SHRI J. S. MONGA SHRI V. J. BHATT (Alternate) (Bombay) SHRI D. N. CHAUDHURI (Alternate) (Calcutta) SHRI L. S. V. EASWAR (Alternate) (Madras) National Test House, Calcutta SHRI S. N. MUKERJI · SHRI B. P. GHOSH (Alternate) All India Instrument Manufacturers' and Dealers' SHRI M. V. NAMJOSHI Association, Bombay SHRI N. BALASUNDARAM (Alternate) Safety for Electronic Equipment Subcommittee. SHRI T. V. RAMAMURTI ETDC 24:3, ISI DR RAM PARSHAD National Physical Laboratory (CSIR), New Delhi Directorate General of Supplies & Disposals (Ministry of Supply, Technical Development SHRI M. SANKARALINGAM & Materials Planning) SHRI P. T. KRISHNAMACHAR (Alternate) SHRI M. V. SHANKAR Railway Board (Ministry of Railways) Directorate General of Posts and Telegraphs SHRI T. S. SUBRAMANIAN (Department of Communications) SHRI S. VENKATARAMAN Radio, Electronics and Television Manufacturers' Association (India), Calcutta SHRI S. C. ISRANI (Aiternate) SHRI Y. S. VENKATESWARAN. Director General, ISI (Ex-officio Member) Director (Elec tech) Secretaries SHRI S. SRINIVASAN Director (Telecom), ISI SHRI N. SRINIVASAN

Deputy Director (Elec tech), ISI

Electronic Measuring Equipment Subcommittee, ETDC 24:2

Convener

KUMARI B. H. SHANTA

Directorate General of Posts and Telegraphs (Department of Communications)

Members

SHRI N. BALASUNDARAM

Eastern Electronics, Faridabad

(Continued on page 8)

Indian Standard

MINIMUM REQUIREMENTS FOR GENERAL PURPOSE AUDIO FREQUENCY SIGNAL GENERATORS

(30 c/s to 30 kc/s)

0. FOREWORD

- 0.1 This Indian Standard was adopted by the Indian Standards Institution on 19 December 1966, after the draft finalized by the Electronic Equipment Sectional Committee had been approved by the Electrotechnical Division Council.
- 0.2 The object of this standard is to specify certain minimum levels of performance expected of general purpose audio frequency signal generators operating in the frequency range of 30 c/s to 30 kc/s intended for servicing of broadcast radio receivers, public address amplifiers, etc.
- **0.2.1** This standard also covers characteristics to be included in the manufacturer's specification for individual instruments and it is anticipated that this standard would ensure a uniform presentation of these characteristics by manufacturers.
- 0.2.2 With regard to safety requirements, the signal generator should satisfy the requirements of IS: 616-1957* in so far as it is applicable to this standard.
- **0.2.3** The signal generators covered by this standard are essentially low power audio frequency voltage generators.
- 0.3 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS: 2-1960†. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

^{*}Code of safety requirements for mains-operated radio receivers.

1. SCOPE

- 1.1 This standard lays down the minimum performance requirements for general purpose audio frequency signal generators operating in the frequency range of 30 c/s to 30 kc/s.
- 1.1.1 Both ac mains-operated and battery-operated signal generators are covered in this standard.
- 1.2 This standard does not cover the climatic, mechnical and physical requirements for audio frequency signal generators.

2. TERMINOLOGY

2.1 For the purpose of this standard, the definitions included in IS: 3915-1966* shall apply.

3. METHODS OF MEASUREMENTS

3.1 The performance requirements specified in this standard are to be measured in accordance with IS: 3915-1966*.

4. OUTPUT SIGNAL FREQUENCY

- 4.1 Frequency Range Frequency range shall generally be 30 c/s to 30 kc/s.
- 4.1.1 The boundary frequencies of each band shall be so adjusted that the upper end of a lower band overlaps with the lower end of the next higher band. The minimum overlap in the boundary frequencies of adjacent bands should be at least 4.0 percent of the highest frequency of lower band to ensure continuous coverage.
- **4.2 Frequency of Internal Calibrator**—The frequency of an internal calibrator, if incorporated, shall correspond to the reference audio frequency of 1000 c/s.
- 4.3 Frequency Calibration Error The basic calibration error at any setting of the frequency shall not exceed 2.0 percent of the scale reading or 3 c/s whichever is greater.
- 4.4 Frequency Drift The frequency drift measured over a period of 1 hour (drift due to long-term operation) shall not exceed 1.0 percent or 1 c/s whichever is greater.
- 4.5 Frequency Change Due to Load The maximum frequency change due to change of load from rated load to half rated load and from rated

^{*}Methods of measurements on audio frequency signal generators (30 c/s to 30 kc/s).

load to open circuit shall not be more than 0.5 percent or 1 c/s whichever is greater.

5. OUTPUT VOLTAGE

5.1 Output Voltage Range — The output voltage range shall be at least 1 mV to 10 V, continuously variable.

Note—This range is applicable when terminated with a load of 600 ohms (see also 6.1).

- 5.2 Output Voltage Variation with Frequency The output voltage shall not vary by more than ± 1 dB over the specified frequency range with respect to the output voltage at 1000 c/s.
- 5.3 Calibrated Voltage The minimum calibrated voltage available shall be at least 10 mV.
- 5.4 Voltage Marking The output voltage markings shall be in millivolts or in volts. Additional marking in dB may also be provided.

Note — In case of dB marking, the reference voltage should correspond to 1 MW through 600 ohms.

- 5.5 Basic Error of Output Voltage The basic error of the output voltage level at any setting in all the bands shall be less than 10 percent.
- 5.6 Hum, Noise and Distortion The total hum, noise and distortion at an output level of 10 V shall not exceed 2.0 percent.

6. SOURCE IMPEDANCE

- **6.1** The source impedance of the signal generator shall be 600 ± 100 ohms.
- 6.1.1 Additional lower impedance (such as 3 ohms) and higher impedance (such as 5000 ohms) may be provided as optional facilities.

7. INFLUENCE OF EXTERNAL OPERATING CONDITIONS

7.1 Influence of Temperature

- 7.1.1 On Frequency The maximum change in the value of frequency over an ambient temperature range of 10° to 50°C shall not exceed 30 percent of the permitted tolerance for frequency calibration error (see 4.3).
- 7.1.2 On Output Voltage The maximum change in the value of output voltage over an ambient temperature range of 10° to 50°C shall

not exceed 30 percent of the permitted tolerance for basic error of the output voltage (see 5.5).

7.2 Influence of Mains Voltage

- 7.2.1 On Frequency—The maximum change in the value of frequency for a mains voltage variation of ± 10 percent shall not exceed 30 percent of the permitted tolerance for frequency calibration error (see 4.3).
- 7.2.2 On Output Voltage The maximum change in the value of output voltage for a mains voltage variation of ± 10 percent shall not exceed 30 percent of the permitted tolerance for basic error of the output voltage (see 5.5).

7.3 Influence of ac Supply Frequency

- 7.3.1 On Frequency The maximum change in the value of frequency for ac supply frequency variation of ± 3 percent from the rated frequency shall not exceed 30 percent of the permitted tolerance for frequency calibration error (see 4.3).
- 7.3.2 On Output Voltage—The maximum change in the value of output voltage for ac supply frequency variation of ± 3 percent from the rated frequency shall not exceed 30 percent of the tolerance for basic error of the output voltage (see 5.5).
- 7.4 Influence of Battery Voltage Under consideration.

8. MARKING

- **8.1** Unless otherwise specified, the following information shall be marked on the signal generator indelibly and externally visible:
 - a) Serial number, and model designation;
 - b) Name and trade mark of the manufacturer; and
 - c) Country of manufacture.
- 8.2 The following information shall be clearly specified in the manual or in the printed label of adequate size affixed to one of the inner faces of the cabinet or back cover:
 - a) Type of supply and the operating voltage,
 - b) Maximum power consumption in the case of mains-operated signal generators or maximum current range in the case of battery-operated signal generators,
 - c) Plan of the chassis showing the location of the components, and
 - d) Any other information or caution which the manufacturer may consider necessary.

8.3 The audio frequency signal generator may also be marked with ISI Certification Mark.

Note — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products as actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

9. TECHNICAL MANUAL

- 9.1 Each signal generator shall be provided with a copy of the technical manual which shall contain the following information:
 - a) Power supply requirements;
 - b) Warm-up period to attain overall stability;
 - c) Source impedance;
 - d) The type of output, namely, balanced or unbalanced; and
 - e) Any other information or caution which the manufacturer may consider necessary.
- 9.1.1 It shall also contain a complete list of parts, spares, and accessories supplied or recommended for use along with a circuit diagram and operating, servicing and maintenance instructions.

(Continued from page 2)

Members

Representing

WG CDB D. B. DAS

Electrical and Electronic Test Measuring Instruments Standardization Panel (Ministry of Defence)

MAJ A. G. DEBAI (Alternate)

SHRI S. K. GUHA

Central Glass and Ceramic Research Institute (CSIR), Calcutta

SHRI S. S. MADAN Dr. G. K. MEHTA Philips India Ltd, Calcutta

Associated Instrument Manufacturers (India)
Pvt Ltd. New Delhi

SHBI N. GANESAN (Alternate)

SHRI J. S. MONGA SHEI R. S. MUNDKUR SHRI D. V. S. RAJU Radiola Corporation, New Delhi Bharat Electronics Ltd, Bangalore

Electronic & Industrial Instruments Co Pvt Ltd, Hyderabad

SHRI V. R. BHAT (Alternate)
DB RAM PARSHAD

National Physical Laboratory (CSIR), New Delhi

SHRI T. N. GHOSH (Alternate) SHRI C. S. RANGAN

National Aeronautical Laboratory (CSIR), Bangalore

SHRI T. V. A. SUBBAMANIAN LT-COL D. SWAROOP Atomic Energy Establishment, Trombay, Bombay Directorate General of Inspection (Ministry of Defence)

SHRIK. V. RAMAMURTHY (Alternate)

SHRI P. K. VENKATASUBRA- Directorate General of All India Radio (Ministry MANIAN of Information & Broadcasting)

INTERNATIONAL SYSTEM OF UNITS (SI UNITS)

QUANTITY	UKIT	SYMBOL	
Length	metre	m	
Mass	kilogram	kg	and the second of the
Time	second	3	
Electric current	ampere	A	
Thermodynamic temperature	kelvin	K	
Luminous intensity	candela	cd	State of the state of the state of
Amount of substance	mole	mol	
Supplementary Units			
QUANTITY	Unit	SYMBOL	
Plane angle	radian	rad	
Solid angle	steradian	sr	
Derived Units			
QUANTITY	UNIT	SYMBOL	DEFINITION
Force	newton	N	1 N = 1 kg.m/s
Energy	joule	J	1J - 1 N.m
Power	watt	W	1 W - 1 J/s
Flux	weber	Wb	1 Wb = 1 V ₄₅
Flux density .	tesla	T	1 T - 1 Wb/m ²
Frequency	hertz	Hr	1 Hz = 1 c/s (s-1)
Electric conductance	siemens	S	1 S = 1 A/V
Electromotive force	volt	V	1 V - 1 W/A
Pressure, stress	pascal	Pa	1 Pa = 1 N/m ²

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002

Telephones: 26 60 21, 27 01 31	Telegrams: Manaksanstha			
Regional Offices:		Telephone		
Western : Novelty Chambers, Grant Road	BOMBAY 400007	6 32 92 95		
Eastern : 5 Chowringhee Approach	CALCUTTA 700072	27 50 90		
Southern : C. I. T. Campus	MADRAS 600113	41 24 42		
Northern : B69, Phase VII	S.A.S. NAGAR (MOHALI) 160051	8 78 26		
Desart Officer	THE RESERVE SHOWS			

AHMADABAD 380001	2 03	3 9	11
BANGALORE 560002	22 48	3 0	15
BHUBANESHWAR 751014	5 3	8 2	27
HYDERABAD 500001	22 10	9 8	13
JAIPUR 302005			
KANPUR 208005	4 7	2 9	12
PATNA 800013	6 2	8 0	18
TRIVANDRUM 695001	33	2 2	27
	BHOPAL 462003 BHUBANESHWAR 751014 HYDERABAD 500001 JAIPUR 302005 KANPUR 208005 PATNA 800013	BANGALORE 560002 22 48 BHOPAL 462003 6 23 BHUBANESHWAR 751014 5 30 HYDERABAD 500001 22 10 JAIPUR 302005 6 98 KANPUR 208005 4 73 PATNA 800013 6 28	BANGALORE 560002 22 48 0 BHOPAL 462003 6 27 1 BHUBANESHWAR 751014 5 36 2 HYDERABAD 500001 22 10 8 JAIPUR 302005 6 98 3 KANPUR 208005 4 72 9 PATNA 800013 6 28 0